

# UNISONICS

## DATA AND SPECIFICATIONS

### **ULTRASONIC CLEANER MODEL FXP10M**

#### **APPLICATIONS:**

The Unisonics FX range is a single chamber device used to clean surgical instruments and other hardware with sonic energy in a heated water / detergent solution. It is designed for use in surgery reprocessing areas, central processing departments and laboratories.

#### **FEATURES:**

- All exterior surfaces re 3mm Palopaque.
- Pressed stainless steel tank.
- A single tank system with rounded corner reflex design.
- Mechanical timer (30 minutes).
- 100 watts of ultrasonic power.
- Polyester coated transducers to prevent moisture contamination and maintain high efficiency levels.
- Internal fitted basket.
- Sealing lid.
- EMC tested and approved.

#### **SONIC CLEANING CHAMBER:** **Free Standing Cabinet**

Overall Size	(External)	Length 263mm Width 162mm Height 240mm
Chamber Size	(Internal)	Length 240mm Width 140mm Depth 100mm
Basket Size		Length 210mm Width 100mm Height 70mm
Maximum Load		5kg
Chamber Volume		3 Litres
Operating Liquid capacity		2 Litres

#### **APPROVALS:**

AS/NZS3760:2001

AS3100-1994-Electrical Safety

Certificate of Conformity No:E990002-(C-Tick)

Certificate for inclusion of medical device-Class 1 (T.G.A.)

# UNISONICS

## DATA AND SPECIFICATIONS

### **ULTRASONIC CLEANER MODEL FXP10M**

#### **CONSTRUCTION:**

All exterior surfaces are 3mm PVC

The ultrasonic chamber is pressed from 304 stainless steel and is 1mm in thickness.

All integral plumbing is stainless steel.

Basket is fabricated from stainless steel mesh and rod.

Sonics energy is provided to the chamber by piezoelectric transducers bonded to the tank bottom with a frequency of 40kHz.

Maximum operating temperature should not exceed 60 degrees Celsius so as to maintain reliability and maximise efficiency.

Minimum operating depth should not be less than 50mm.

Power is supplied by solid-state circuitry, which is air-cooled.

#### **CONTROLS:**

A 30-minute mechanical timer control is located at the bottom front of the unit and should be set as per operating instructions.

The mains plug/socket is located at the side of the unit.

The unit should always be positioned to allow for plenty of circulation.

Care should always be taken to avoid excessive spillage of solution when draining fluid.

#### **GENERAL CLEANING:**

The turnover of solution must be determined by the user to satisfy acceptable cleaning results of the finished article.

The more solution is contaminated the longer the cleaning process.

If the contamination is heavy and difficult to remove a pre rinse in a heated bath could be sufficient enough to soften and loosen unwanted debris in preparation for the ultrasonic cleaning.

At the other end of the scale after the items have been removed it is essential that a final wash be performed to remove any residue that remains on the cleaned parts.