

UNISONICS

DATA AND SPECIFICATIONS

ULTRASONIC CLEANER MODEL FXP40D

APPLICATIONS:

The Unisonic FX range is a single chamber device used to clean surgical instruments and other hardware with sonic energy in a heated water/detergent solution. It is designed for use in surgery reprocessing areas, central processing departments and laboratories.

FEATURES:

- All exterior surfaces are 6mm Palopaque.
- Pressed stainless steel tank.
- A single tank system with rounded corner reflex.
- Digital timer.
- 1200 watts of ultrasonic power.
- Polyester coated transducers to prevent moisture contamination and maintain high efficiency levels.
- Easy access drain valve.
- Internal fitted basket.
- Sealing lid.
- Cooling fan.
- Outlet drain.
- EMC tested and approved.

SONIC CLEANING CHAMBER

Free standing cabinet

Overall Size	(External)	Length 640mm Width 530mm Height 400mm
Chamber Size	(Internal)	Length 600mm Width 500mm Depth 200mm
Basket Size		Length 550mm Width 450mm Depth 140mm
Maximum Load		10kg
Chamber Volume		60 litres
Operating liquid capacity		45 litres

APPROVALS:

AS/NZS3760:2001

AS3100-1994-Electrical Safety

Certificate of Conformity No:E990002-(C-Tick)

Certificate for inclusion of medical device-Class 1 (T.G.A.)

UNISONICS

DATA AND SPECIFICATIONS

ULTRASONIC CLEANER MODEL FXP40D

CONSTRUCTION:

All exterior surfaces are 6mm PVC.

The Ultrasonic chamber is pressed from 304 stainless steel and is 1mm in thickness.

All integral plumbing is stainless steel.

Two fans are located on each side to allow an efficient cross flow of cool air.

A valve-draining outlet is located on the left hand side of the unit away from electrical connections.

Basket is fabricated from stainless mesh and rod.

Sonic energy is provided to the chamber by piezoelectric transducers bonded to the tank bottom with a frequency of 40kHz.

Maximum operating temperature should not exceed 60 degrees Celsius so as to maintain reliability and maximise efficiency.

Minimum operating depth should not be less than 80mm.

Power is supplied by solid-state circuitry, which is air-cooled.

CONTROLS:

A digital timer control is located at the bottom front of the unit and should be set as per operating instructions.

The mains plug/socket outlet is located at the side of the unit, which allows the operator to remove the connection when the cleaner is not in use.

The unit should always be positioned to allow for plenty of circulation.

Care should always be taken to avoid excessive spillage of solution when draining fluid.

GENERAL CLEANING:

The turnover of solution must be determined by the user to satisfy acceptable cleaning results of the finished article.

The more the solution is contaminated the longer the cleaning process.

If the contamination is heavy and difficult to remove a pre-rinse in a heated bath could be sufficient enough to soften and loosen unwanted debris in preparation for the ultrasonic cleaning.

At the other end of the scale after the items have been removed it is essential that a final wash be performed to remove any residue that remains on the cleaned parts.